IEC 61526

Edition 4.1 2026-06

INTERNATIONAL STANDARD

CONSOLIDATED VERSION

Radiation protection instrumentation - Measurement of personal dose equivalents for X, gamma, neutron and beta radiations - Active personal dosemeters

ICS 13.280

ISBN 978-2-8327-0000-0

IEC 61526:2024-03+AMD1:2026-06 CSV(en)

CONTENTS

FOREWORD 6
INTRODUCTION 8
1 Scope 9
2 Normative references 9
3 Terms and definitions 12
4 Units and symbols 19
4.1 Units 19
4.2 Symbols 19
5 Mechanical characteristics 21
5.1 Case 21
5.2 Switches 22
6 General characteristics 22
6.1 Storage of dose information 22
6.2 Indication 22
6.3 Dosemeter markings 22
6.4 Retention of radioactive contamination 22
6.5 Ranges for dose and dose rate 22
6.6 Rated range of an influence quantity 22
6.7 Effect of radiation not intended to be measured and use of more than one dosemeter 23
6.8 Intrinsic and environmental background of the instrument 23
6.9 Dose or dose rate alarms 23
6.9.1 General 23
6.9.2 Dose alarms 23
6.9.3 Dose rate alarms 24
6.9.4 Alarm output 24
6.10 Indication of malfunction 24
6.11 Assignment of the dose value to the dosemeter 24
6.12 Algorithm to evaluate the indicated value 24
7 General test procedures 25
7.1 Nature of tests 25
7.2 Reference conditions and standard test conditions 25
7.3 Tests for influence quantities of type F 26
7.4 Tests for influence quantities of type S 26
7.5 Phantom for testing 26
7.6 Position of detector assembly for the purpose of testing 26
7.7 Position of dosemeter during use 27
7.8 Minimum rated range of influence quantity 27
7.9 Low dose rates 27
7.10 Statistical fluctuations 27
7.11 Production of reference radiation 27
8 Additivity of indicated value 27
8.1 Requirements 27
8.2 Method of test 28
8.2.1 General 28
8.2.2 Preparation of the test 28

8.2.3 Practical test 29
8.3 Interpretation of the results 29
9 Radiation performance requirements and tests 29
9.1 General 29
9.2 Consideration of the uncertainty of the conventional quantity value 30
9.3 Constancy of the dose response, dose rate dependence and statistical fluctuations 30
9.3.1 General 30
9.3.2 Requirements 30
9.3.3 Method of test using sources 35
9.3.4 Interpretation of the results of the test using sources 36
9.3.5 Method of test for photon dosemeters using natural radiation 37
9.3.6 Interpretation of the results of the test using natural radiation 37
9.4 Variation of the response due to photon radiation energy and angle of incidence 38
9.4.1 Measurement quantity $H_p(0,07)$ or $\dot{H}_p(0,07)$ 38
9.4.2 Measurement quantity $H_p(3)$ or $\dot{H}_p(3)$ 38
9.4.3 Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$ 39
9.5 Variation of the response due to neutron radiation energy and angle of incidence 40
9.5.1 General 40
9.5.2 Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$ 40
9.6 Variation of the response due to beta radiation energy and angle of incidence 41
9.6.1 General 41
9.6.2 Measurement quantity $H_p(0,07)$ or $\dot{H}_p(0,07)$ 41
9.6.3 Measurement quantity $H_p(3)$ or $\dot{H}_p(3)$ 42
9.6.4 Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$ 43
9.7 Over indication due to radiation incident from the side of an $H_p(10)$, $H_p(3)$ or $H_p(0,07)$ dosemeter 43
9.7.1 Requirements 43
9.7.2 Method of test 43
9.7.3 Interpretation of the results 44
9.8 Retention of dose reading 44
9.8.1 General 44
9.8.2 Requirements 44
9.8.3 Method of test and interpretation of the results 44
9.9 Overload characteristics 44
9.9.1 General 44
9.9.2 Requirements 45
9.9.3 Method of test and interpretation of the results 45
9.10 Alarm 46
9.10.1 General 46
9.10.2 Response time for dose rate indication and alarm 46
9.10.3 Accuracy of dose alarm 47
9.10.4 Accuracy of dose rate alarm 47
9.11 Model function 48

10 Electrical and environmental performance requirements and tests 48
10.1 General 48
10.2 Power supplies 48
10.2.1 General requirements 48
10.2.2 Specific primary batteries requirements 49
10.2.3 Specific secondary batteries requirements 49
10.2.4 Method of test and interpretation of the results (primary and secondary batteries) 49
10.3 Ambient temperature 50
10.3.1 Requirements 50
10.3.2 Method of test and interpretation of the results 51
10.4 Relative humidity 52
10.4.1 Requirements 52
10.4.2 Method of test and interpretation of the results 52
10.5 Atmospheric pressure 52
10.6 Sealing 52
10.7 Storage 52
10.8 Light exposure 52
10.8.1 General 52
10.8.2 Requirements 53
10.8.3 Interpretation of the results 53
10.9 Dose build-up, fading and self-irradiation (hybrid dosemeters) 53
11 Electromagnetic performance requirements and tests 54
11.1 General 54
11.2 Electrostatic discharge 55
11.2.1 Requirements 55
11.2.2 Method of test and interpretation of the results 55
11.3 Radiated electromagnetic fields 56
11.3.1 Requirements 56
11.3.2 Method of test and interpretation of the results 56
11.4 Conducted disturbances induced by fast transients or bursts 56
11.4.1 Requirements 56
11.4.2 Method of test and interpretation of the results 56
11.5 Conducted disturbances induced by surges 57
11.5.1 Requirements 57
11.5.2 Method of test and interpretation of the results 57
11.6 Conducted disturbances induced by radio-frequencies 57
11.6.1 Requirements 57
11.6.2 Method of test and interpretation of the results 57
11.7 50 Hz/60 Hz magnetic field 57
11.7.1 Requirements 57
11.7.2 Method of test and interpretation of the results 57
11.8 Voltage dips and short interruptions 57
11.8.1 Requirements 57
11.8.2 Method of test and interpretation of the results 58
12 Mechanical performance, requirements and tests 58
12.1 General 58
12.2 Drop test 58

12.2.1 Requirements 58
12.2.2 Method of test and interpretation of the results 58
12.3 Vibration test 59
12.3.1 Requirements 59
12.3.2 Method of test and interpretation of the results 59
12.4 Microphonics test 59
12.4.1 Requirements 59
12.4.2 Method of test and interpretation of the results 59
13 Software 60
13.1 General 60
13.2 Design and structure of the software 61
13.2.1 Requirements 61
13.2.2 Method of test 61
13.3 Identification of the software 61
13.3.1 Requirements 61
13.3.2 Method of test 61
13.4 Authenticity of the software and the presentation of results 62
13.4.1 Requirements 62
13.4.2 Method of test 62
13.5 Alarm and stop of system operation under abnormal operating conditions 62
13.5.1 Requirements 62
13.5.2 Method of test 62
13.6 Control of instrument parameters 63
13.6.1 Requirements 63
13.6.2 Method of test 63
13.7 Storage of measurement data 63
13.7.1 Requirements 63
13.7.2 Method of test 63
13.8 Transmission of data 64
13.8.1 Requirements 64
13.8.2 Method of test 64
13.9 Hardware interfaces and software interfaces 64
13.9.1 Requirements 64
13.9.2 Method of test 65
13.10 Documentation for the software test 65
13.10.1 Requirements 65
13.10.2 Method of test 66
14 Uncertainty 66
15 Documentation 66
15.1 Type test report 66
15.2 Certificate 66
16 Operation and maintenance manual 67
Annex A (normative) Statistical fluctuations 68
~~Annex B (normative) Coverage intervals ..... 73~~
Annex C (informative) Causal connection between readout signals, indicated value and measured value 73
Annex D (informative) Procedure to determine the variation of the relative response due to radiation energy and angle of radiation incidence 74

Annex E (informative) Computational method of test for mixed irradiations 76
Bibliography 78

Figure B.1 – Test for coverage interval 70
Figure C.1 – Data evaluation in dosimetry systems 73
Figure E.1 – Flow chart of a computer program to perform tests according to 8.2 77

Table 1 – Symbols and abbreviated terms 19
Table 2 – Reference conditions and standard test conditions 25
Table 3 – Radiation characteristics of $H_p(0,07)$ dosemeters for X, gamma and beta radiation 31
Table 4 – Radiation characteristics of $H_p(3)$ dosemeters for X, gamma and beta radiation 32
Table 5 – Radiation characteristics of $H_p(10)$ dosemeters for X and gamma radiation 33
Table 6 – Radiation characteristics of $H_p(10)$ dosemeters for neutron radiation 34
Table 7 – Values of c_1 and c_2 for w different dose values and n indications for each dose value 37
Table 8 – Electrical and environmental characteristics of dosemeters 53
Table 9 – Electromagnetic disturbance characteristics of dosemeters 55
Table 10 – Mechanical disturbances characteristics of dosemeters 58
Table A.1 – Number of instrument readings required to detect true differences (95 % level of confidence) between two sets of instrument readings on the same instrument 69
Table B.1 – Student's t-value for a double sided 95 % coverage interval 71
Table E.1 – Example of dosemeter response table and range limits 76

INTERNATIONAL ELECTROTECHNICAL COMMISSION

_____

Radiation protection instrumentation - Measurement of personal dose equivalents for X, gamma, neutron and beta radiations - Active personal dosemeters

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at https://patents.iec.ch. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 61526 edition 4.1 contains the fourth edition (2024-03) [documents 45B/1047/FDIS and 45B/1049/RVD] and its amendment 1 (2026-06) [documents 45B/1113/FDIS and 45B/1121/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

IEC 61526 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation. It is an International Standard.

This fourth edition cancels and replaces the third edition published in 2010. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

a) Modification of the title;
b) Inclusion of the measurement quantity for the dose in the lens of the eye, $H_p(3)$;
c) Inclusion of measurement quantity for dose in the skin and extremities, $H_p(0,07)$;
d) Inclusion of dosemeters between active and passive: "hybrid dosemeters";
e) Inclusion of software requirements;
f) Harmonization of requirements for linearity to IEC 62387;
g) Revised neutron energy response requirements.

The text of this International Standard is based on the following documents:

Draft	Report on voting
45B/1047/FDIS	45B/1049/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

– reconfirmed,
– withdrawn, or
– revised.

INTRODUCTION

This document applies to active, (direct reading or hybrid) personal dosemeters and monitors used for measuring personal dose equivalents $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, for X, gamma, neutron, and beta radiations.

For personal dose equivalent $H_p(10)$ and for X and gamma radiation, two minimum rated ranges for the photon energy are given. The first from 20 keV to 150 keV is for workplaces where low energy X-rays are used, e.g., in diagnostic medicine, the second from 80 keV to 1,25 MeV is for workplaces where high energy X-rays and/or gamma sources are used, e.g., in industry. For neutron radiation the minimum rated range of neutron energy is from 0,025 eV (thermal neutrons) to 10 MeV. The rated ranges can be extended to all energies covered by the respective standards for reference radiation fields.

For personal dose equivalent $H_p(3)$ and for X and gamma radiation, a minimum rated range for photon energy from 30 keV to 250 keV is given. For personal dose equivalent $H_p(0,07)$ a range of 30 keV to 1250 keV or, for workplaces where low energy X-rays are used, 20 keV to 150 keV, is given. For beta radiation for both quantities, the minimal rated range is from 0,24 MeV to 0,8 MeV (mean beta particle energy). The rated ranges can be extended to all energies covered by the respective standards for reference radiation fields.

In some applications, for example, at a nuclear reactor installation where 6 MeV photon radiation is present, measurement of personal dose equivalent (rate) $H_p(10)$ for photon energies up to 10 MeV should be required. In some other applications, measurement of $H_p(10)$ down to 10 keV should be required.

For personal dosemeters, requirements for measuring the dose quantities $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, and for monitoring of the respective dose rate quantities are given. The measurement of these dose rate quantities is an option for personal dosemeters.

Establishments in some countries may be permitted to use this type of personal dosemeter as the dosemeter to provide the dose of record by an approved dosimetry service.

1 Scope

This document applies to personal dosemeters with the following characteristics:

a) They are worn on the trunk, close to the eye, or on the extremities.
b) They measure the personal dose equivalents $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, from external X and gamma, neutron (not for $H_p(3)$), and beta radiations, and may measure the respective personal dose equivalent rates for the same radiations (for alarming purposes).
c) They have a digital indication. This indication may or may not be attached.
d) They have alarm functions for the personal dose equivalents or personal dose equivalent rates except for hybrid dosemeters. For hybrid dosemeters an alarm function for the personal dose equivalents shall be implemented in the associated readout system.

NOTE 1 When reference is made in this document to "dose", this is meant to indicate personal dose equivalent, unless otherwise stated.

NOTE 2 When reference is made in this document to "dosemeter", this is meant to include all personal dosemeters, unless otherwise stated.

NOTE 3 This document does not cover neutron dosemeters for the $H_p(3)$ measurements up to now.

This document specifies requirements for the dosemeter and, if supplied, for its associated readout system.

Usually, a dosemeter is not able to measure all quantities given above. Thus, the dosemeter is only tested with regard to those quantities and types of radiation it is intended to be used for. Other types of radiation are considered as influence quantities which also may have requirements.

This document specifies, for the dosemeters described above, general characteristics, general test procedures, radiation characteristics as well as electrical, mechanical, safety and environmental characteristics. The only requirements specified for associated readout systems are those which affect its accuracy of readout of the personal dose equivalent and alarm settings, and those which concern the influence of the reader on the dosemeter.

This document does not cover special requirements for accident or emergency dosimetry, although the dosemeters may be used for this purpose.

This document does not apply to dosemeters used for measurement of pulsed radiation, such as radiation emanating from many medical diagnostic X-ray facilities, linear accelerators or similar equipment.

NOTE 4 Requirements and testing procedure for dosemeters used in pulsed field of ionizing radiation can be found in IEC TS 63050 or IEC TS 62743.

This document does not apply for dosemeters to measure ambient or directional dose equivalent.

NOTE 5 Requirements for ambient or directional dose equivalent meters can be found e.g. in: IEC 60846-1; IEC 62387, IEC 61017 or IEC 60532.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-395:2014, *International Electrotechnical Vocabulary (IEV) – Part 395: Nuclear instrumentation – Physical phenomena, basic concepts, instruments, systems, equipment and detectors*
IEC 60050-395:2014/AMD1:2016
IEC 60050-395:2014/AMD2:2020

IEC 60068-2-31:2008, *Environmental testing – Part 2-31: Tests – Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60086-1:2021, *Primary batteries – Part 1: General*

IEC 60086-2:2021, *Primary batteries – Part 2: Physical and electrical specifications*

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*
IEC 60529:1989/AMD1:1999
IEC 60529:1989/AMD2:2013

IEC 60904-3, *Photovoltaic devices – Part 3: Measurement principles for terrestrial photovoltaic (PV) solar devices with reference spectral irradiance data*

IEC 61000-4-2, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-4-3, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-4, *Electromagnetic compatibility (EMC) – Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test*

IEC 61000-4-5, *Electromagnetic compatibility (EMC) – Part 4-5: Testing and measurement techniques – Surge immunity test*

IEC 61000-4-6, *Electromagnetic compatibility (EMC) – Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-4-8, *Electromagnetic compatibility (EMC) – Part 4-8: Testing and measurement techniques – Power frequency magnetic field immunity test*

IEC 61000-4-11, *Electromagnetic compatibility (EMC) – Part 4-11: Testing and measurement techniques – Voltage dips, short interruptions and voltage variations immunity tests for equipment with input current up to 16 A per phase*

IEC 61000-6-2, *Electromagnetic compatibility (EMC) – Part 6-2: Generic standards – Immunity for industrial environments*

IEC 61187:1993, *Electrical and electronic measuring equipment – Documentation*

IEC 62387:2020, *Radiation protection instrumentation – Dosimetry systems with integrating passive detectors for individual, workplace and environmental monitoring of photon and beta radiation*

IEC TR 62461:2015, *Radiation protection instrumentation – Determination of uncertainty in measurement*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement* (GUM:1995)

ISO/IEC Guide 98-3:2008/Suppl.1:2008, *Propagation of distributions using a Monte Carlo method and Corr.1 (2009)*

ISO 4037-1:2019, *Radiological protection – X and gamma reference radiation for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy – Part 1: Radiation characteristics and production methods*

ISO 4037-2:2019, *Radiological protection – X and gamma reference radiation for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy – Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV*

ISO 4037-3:2019, *Radiological protection – X and gamma reference radiation for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy – Part 3: Calibration of area and personal dosemeters and the measurement of their response as a function of energy and angle of incidence*

ISO 4037-4:2019, *Radiological protection – X and gamma reference radiation for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy – Part 4: Calibration of area and personal dosemeters in low energy X reference radiation fields*

ISO 6980-1:2023, *Nuclear energy – Reference beta-particle radiation – Part 1: Methods of production*

ISO 6980-2:2023, *Nuclear energy – Reference beta-particle radiation – Part 2: Calibration fundamentals related to basic quantities characterizing the radiation field*

ISO 6980-3:2023, *Nuclear energy – Reference beta-particle radiation –Part 3: Calibration of area and personal dosemeters and the determination of their response as a function of beta radiation energy and angle of incidence*

ISO 8529-1:2021, *Reference neutron radiations – Part 1: Characteristics and methods of production*

ISO 8529-2:2000, *Reference neutron radiations – Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field*

ISO 8529-3:2023, *Neutron reference radiation fields – Part 3: Calibration of area and personal dosemeters and determination of their response as a function of neutron energy and angle of incidence*

ISO 12789-1, *Reference radiation fields – Simulated workplace neutron fields – Part 1: Characteristics and methods of production*

ISO 12789-2, *Reference radiation fields – Simulated workplace neutron fields – Part 2: Calibration fundamentals related to the basic quantities*

ISO 21909-1:2021, *Passive neutron dosimetry systems – Part 1: Performance and test requirements for personal dosimetry*

ISO 80000-10:2019, *Quantities and units – Part 10: Atomic and nuclear physics*

ICRU Report 51:1993, *Quantities and units in radiation protection dosimetry*

ANSI N42.17A, *American National Standard For Performance Specifications For Health Physics Instrumentation – Portable Instrumentation For Use In Normal Environmental Conditions*

Bibliography

IEC 60050-151:2001, *International Electrotechnical Vocabulary (IEV) – Part 151: Electrical and magnetic devices*
IEC 60050-151:2001/AMD1:2013
IEC 60050-151:2001/AMD2:2014
IEC 60050-151:2001/AMD3:2019
IEC 60050-151:2001/AMD4:2020
IEC 60050-151:2001/AMD5:2021

IEC 60050-300:2001, *International Electrotechnical Vocabulary (IEV) – Part 300: Electrical and electronic measurements and measuring instruments – Part 311: General terms relating to measurements – Part 312: General terms relating to electrical measurements – Part 313: Types of electrical measuring instruments – Part 314: Specific terms according to the type of instrument*
IEC 60050-300:2001/AMD1:2015
IEC 60050-300:2001/AMD2:2016
IEC 60050-300:2001/AMD3:2017
IEC 60050-300:2001/AMD4:2020

IEC 60532, *Radiation protection instrumentation – Installed dose rate meters, warning assemblies and monitors – X and gamma radiation of energy between 50 keV and 7 MeV*

IEC 60846-1, *Radiation protection instrumentation – Ambient and/or directional dose equivalent (rate) meters and/or monitors for beta, X and gamma radiation – Part 1: Portable workplace and environmental meters and monitors*

IEC 61017, *Radiation protection instrumentation – Transportable, mobile or installed equipment to measure photon radiation for environmental monitoring*

IEC 62706, *Radiation protection instrumentation – Recommended climatic, electromagnetic and mechanical performance requirements and methods of tests*

IEC TS 62743:2012, *Radiation protection instrumentation – Electronic counting dosemeters for pulsed fields of ionizing radiation*

IEC TS 63050, *Radiation protection instrumentation – Dosemeters for pulsed fields of ionizing radiation*

ISO/IEC Guide 99:2007, *International vocabulary of metrology – Basic and general concepts and associated terms (VIM)*

ISO 29661:2012, *Reference radiation fields for radiation protection – Definitions and fundamental concepts*

Brunzendorf, J. and Behrens, R., "How to type test the coefficient of variation of an indication", *Radiation Protection Dosimetry*, Vol. 123, pp. 21-31,2007

WELMEC European cooperation in legal metrology, WELMEC 7.2: Software Guide (Measuring Instruments Directive 2014/32/EU), available for download at www.welmec.org

_____